


Ochratoxin

A and B

Romer Labs® clean-up columns for Ochratoxin offer the versatility in analysis depending on the complexity of the commodity, the choice of detection method used and the speed one desires to spend on sample clean-up.



1-step clean up

MycoSep®229 Ochra



- Time for clean up: <60 seconds
- 2 formats of the same column
- Stability: 18 months
- Recoveries: >95%
- Repeatability: <5%
- For cereals, grains, raisins, green coffee, red wine...

MultiSep®229 Ochra




3-step clean up

OchraStar™ column




- Compliant with EC commission directive 2002/26/EC
- Recoveries: >95%
- Repeatability: <5%
- Different extraction solutions can be used
- Stability: 18 months
- For complex commodities and foodstuffs

Procedure for Ochratoxin A and B Clean up




1-step clean up

**MycoSep[®] 229
Ochra
principle**



1-step clean up

**MultiSep[®] 229
Ochra
principle**



3-step clean up

**OchraStar[™]
principle**

Order Information:

MycoSep[®]229 Ochra column
Order No.:COCMY2229, Box of 25
MultiSep[®]229 Ochra column
Order No.:COCMU2229, Box of 25

OchraStar[™] column
Order No.:COIAC2000, Box of 25

Standard:
Ochratoxin A
Order No.:002023